

TLN

**Department of Forensic Science
Breath Alcohol Section
Instrument Location History
1/1/1990 To 6/12/2014**

Instrument Serial Number: 010443

As of 12-Jun-14

Start Date	End Date	Agency Name
12-Jun-14		DFS Central Lab
24-Aug-12	12-Jun-14	Prince George County PD
25-Jul-12	24-Aug-12	DFS Central Lab
16-Jun-09	25-Jul-12	Roanoke County Sheriff's Office
22-Apr-09	16-Jun-09	DFS Central Lab
06-Nov-08	22-Apr-09	South Hill PD
13-May-08	06-Nov-08	DFS Central Lab

TLN

INTOX EC/IR II **Quality Assurance Worksheet**

Instrument Serial Number **010443** Worksheet Start Date **6/12/2014**

Location **Prince George PD**

Address **6600 Courthouse Rd Prince George VA 23875**

DFS Technician **Timmy Neece, Jr** License No. **25387**

☐ Laboratory ☒ On-Site

Site Specification: No detrimental environmental conditions exist. ☒

Instrument Barometer (mm HG) **755** Reference Barometer (mm HG) **757**

Reference Barometer(RB)Serial # **009110** RB Calibration Due **5/28/2015**

Measurement Assurance Check

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.299
0.300	0.298	0.289	0.306	Sample 2	0.299
Precision		sample min	sample max	Sample 3	0.298
0.001		0.298	0.299		

Standard (sea level)	PA Target	minimum	maximum	Sample 1	0.100
0.100	0.099	0.096	0.102	Sample 2	0.100
Precision		sample min	sample max	Sample 3	0.100
0		0.100	0.100		


Dry gas standard Lot No. (with tank no.) **AG227501-04**

☐ Replaced dry gas standard (+O-ring)

☐ Installed at Location

☒ Removed to DFS-Central

Supplies
Mouthpieces
Certificates of Analysis
Operator Worksheet
Other:



Notes:

Instrument Serial Number

010443

Certification Date

☐ Calibrated☐ Certified☒ Measurement Assurance Check☒ Instrument Test☐ Troubleshooting/Maintenance Worksheet Completed

(Successful completion denotes satisfactory condition of the item.)

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.250		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.150		0.000	0.000	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	


Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.080		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

Standard (sea level)	PA Target	minimum	maximum	Sample 1
0.020		-0.003	0.003	Sample 2
Precision		sample min	sample max	Sample 3
0		0.000	0.000	

All measurements are in g/210L

Estimation of Uncertainty of Measurement and traceability records are located within the Breath Alcohol Section


DFS Technician



Date

6/12/14

Issuing Analyst



Date

6/12/14



TLN




COMMONWEALTH OF VIRGINIA
DEPARTMENT OF FORENSIC SCIENCE

CERTIFICATE OF BLOOD ALCOHOL ANALYSIS
AS DETERMINED BY A CHEMICAL TEST OF THE ACCUSED'S BREATH

NAME OF ACCUSED INSTRUMENT, TEST,	NAME OF COURT DFS
---	-----------------------------

BREATH ANALYSIS

SAMPLE EXAMINED AND TEST CONDUCTED BY NEECE, JR., TIMMY, L		AGENCY DFS Central Lab
DFS LICENSE NUMBER 25387	LICENSE EXPIRES 10/01/2014	DATE TEST CONDUCTED 06/12/2014
TEST EQUIPMENT NUMBER 010443		

RESULTS: TIME SAMPLE TAKEN 10:18 EDT

SAMPLE'S ALCOHOL CONTENT 0.00 GRAMS PER 210 LITERS OF BREATH

ATTEST:

I CERTIFY THAT THE ABOVE IS AN ACCURATE RECORD OF THE TEST CONDUCTED; THAT THE TEST WAS CONDUCTED WITH THE TYPE OF EQUIPMENT AND IN ACCORDANCE WITH THE METHODS APPROVED BY THE DEPARTMENT OF FORENSIC SCIENCE; THAT THE TEST WAS CONDUCTED IN ACCORDANCE WITH THE DEPARTMENT'S SPECIFICATIONS; THAT PRIOR TO ADMINISTRATION OF THE TEST THE ACCUSED WAS ADVISED OF HIS RIGHT TO OBSERVE THE PROCESS AND SEE THE BLOOD ALCOHOL READING ON THE EQUIPMENT USED TO PERFORM THE BREATH TEST, AND THAT I POSSESS A VALID LICENSE TO CONDUCT SUCH TEST, GIVEN UNDER MY HAND THIS _____ DAY OF _____, 20____.

BREATH TEST OPERATOR

☐ I HAVE RECEIVED A COPY OF THIS CERTIFICATE OF ANALYSIS _____ SUBJECT'S SIGNATURE

☐ SUBJECT REFUSED TO SIGN FOR COPY OF CERTIFICATE OF ANALYSIS _____ OPERATOR'S SIGNATURE

FOR COURT



IntoxNet MIS Report

Report Generated 12 Jun 2014 at 10:47

HMS
TLN

Test Results

Instrument Serial Number 010443

Test # 001791 Subject Test

Test Location 1 Department of
Test Date 12 Jun 2014

Test Location 2 Forensic Science
Test Time 10:12
Remote/Local Local

Test Location 3
System Check Passed

Operator's Last Name NEECE, JR.
Agency DFS Central Lab

Operator's First Name TIMMY

Operator's Middle Initial L

Card Serial Number 125387

Effective Date 10/01/2012

License Number 25387

Expiration Date 10/01/2014

Subject's Last Name INSTRUMENT

Subject's First Name TEST

Subject's Middle Initial

Subject's Date of Birth 00/00/0000

Subject's Sex Male

Driver's License Number

Driver's License Expiration 00/00/0000

Driver's License State

Court Name DFS

End Date 12 Jun 2014 End Time 10:20

Result Time 10:18

Result Date 12 Jun 2014 Result 0.00

Data Type DIAG

Sample Value Pass

Sample Time 10:12

Data Type BLK

Sample Value 0.000

Sample Time 10:13

Data Type CHK

Sample Value 0.100

Sample Time 10:14

Data Type BLK

Sample Value 0.000

Sample Time 10:15

Data Type SUBJ

Sample Value 0.000

Sample Time 10:16

Data Type BLK

Sample Value 0.000

Sample Time 10:17

Data Type SUBJ

Sample Value 0.000

Sample Time 10:18

Data Type BLK

Sample Value 0.000

Sample Time 10:19

Standard Type Dry Gas Std

Standard Value 0.099

Standard Lot Number AG227501-04

Standard Expiration Date 10/01/2014

Tank Pressure 237

Barometric Pressure 755 mmHg

Blow Sample Number 1 Blow Duration 3.11 sec

Blow Volume 1841 cc End-of-Blow Time 10:16

Blow Sample Number 2 Blow Duration 4.51 sec

Blow Volume 2090 cc End-of-Blow Time 10:18

Tamper Evident Stamp 64aaf68d

Test Status Code 0

Test Status Success

TL✓

INTOX EC/IR-11: Accuracy Check

Department of Forensic Science

Serial Number: 010443 Test Number: 1792

Test Date: 06/12/2014 Test Time: 10:21 EDT

Dry Gas Target: 0.099

Lot Number: AG227501-04 Exp Date: 10/01/2014

Tank Pressure: 232 psi Barometric Pressure: 755 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:22
CHK	0.100	10:22
BLK	0.000	10:24
CHK	0.100	10:25
BLK	0.000	10:26
CHK	0.100	10:27

Test Status: *Success*

Calibration CRC: E14FD470

TLN

Intox EC/1R-11: Accuracy Check

Department of Forensic Science

Serial Number: 010443 Test Number: 1793

Test Date: 06/12/2014 Test Time: 10:30 EDT

Dry Gas Target: 0.298

Lot Number: AG334603-07 Exp Date: 12/12/2015

Tank Pressure: 272 psi Barometric Pressure: 755 mmHg

System Check: *Passed*

Test	g/210L	Time
BLK	0.000	10:31
CHK	0.299	10:31
BLK	0.000	10:33
CHK	0.299	10:34
BLK	0.000	10:35
CHK	0.298	10:36

Test Status: *Success*

Calibration CRC: E14FD470